



Attestation of Conformity

No. ICR Polska/ M3600414



| | |
|---|---|
| Name and address of Registered Manufacturer: | Guangxi WBS Medical Technology Co., Ltd. No.2 of Zhengrun Boulevard of Eco-Industrial Park, Hezhou, Guangxi |
| Product name: | Infrared Thermometer |
| Product type/model: | WBS-T003, WBS-T007, WBS-T008 |
| Rating: | 3VDC, 1A |
| Trade mark: | n.a. |

This Attestation confirms that the product meets the requirements of the following normative documents and within limits of its documents gives presumption of conformity with essential requirements of Directive 93/42/EEC.

| | |
|--|---|
| Relevant EC Directive: | Medical Device Directive 93/42/EEC |
| Conformity assessment procedure: | EC Declaration of Conformity (Annex VII of Directive 93/42/EEC) |
| Classification: | Class I according Rule 1 of Annex IX of Directive 93/42/EEC |
| Applied normative documents: | EN 60601-1:2006+A1:2013+A12:2014 EN 60601-1-2: 2015 |
| Applied Quality Management System | n/a |

This AoC will remain valid only if Quality Management System Certificate remains valid.
The assessment process has been carried out in accordance with the program PC-P-07-07.
Evaluation has been carried out in accordance with test report made by:

- Shenzhen Anbotek Compliance Laboratory Limited

| | |
|----------------------------|------------------------------------|
| No. of test report: | 18250IC00001701E, 18250IC00001701S |
| Issue date: | 25.03.2020 |
| Expiration date: | 24.03.2025 |

The mutual obligations and rights of the certification are regulated by the contract No. ICR Polska/2020-3125.

This Attestation applies to products having the same attributes (parameters), intended use, that have been evaluated and meet the requirements of the aforementioned standard.



Director: Rafał Kalinowski

Warsaw, 25. 03. 2020.

ICR Polska Co. Ltd.

ul. Plac Przymierza 6, 03-944 Warszawa
www.icrpolska.com, e-mail: icrpolska@icrq.com

